

### 湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

地址 (Add): 湖北省武汉市东湖新技术开发区茅店山中路二号  
邮编 (Post Code): 430223  
电话 (Tel): 027-81925136  
网址 (Web site): <http://www.hbjl.gov.cn>  
传真 (Fax): 027-81925137

中国校准服务联合体  
China United Calibration Service

**CUC**

REGISTRATION NO. 002


# 校准证书

CALIBRATION CERTIFICATE

证书编号: [2020DW04400243]  
Certificate No.

|                        |                |
|------------------------|----------------|
| 委托方名称<br>Customer      | 安徽丰原热电有限公司     |
| 委托方地址<br>Address       | 安徽省蚌埠市         |
| 样品名称<br>Name Of Sample | 绝缘油介电强度测试仪     |
| 制造厂商<br>Manufacturer   | 武汉三新电力设备制造有限公司 |
| 型号规格<br>Model/Type     | SXIJJ-III      |
| 器具编号<br>No Of Sample   | 11904038       |

证书专用章  
Stamp



校准日期  
Calibration date

2020 年 06 月 23 日  
Y M D

证书批准人  
Approved by

耿睿

核验员  
Checked by

李亮波

校准员  
Calibrated by

董妙宇

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。  
The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



湖北省计量测试技术研究院  
Hubei Institute of Measurement and Testing Technology

证书编号: [2020DW04400243]号  
Certificate No.

- 本院是政府计量行政管理部门依法设立的法定计量检定机构  
This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.
- 本院质量管理体系符合ISO/IEC17025标准的要求。  
The quality management system for laboratory complies with ISO/IEC 17025 standards.
- 本次校准的技术依据（名称、代号）  
Reference documents for the Calibration (Name、Code)  
参照: JJG 496-2016 工频高压分压器

- 本次校准所使用的主要计量标准器具  
Main standards of measurement used in the Calibration

| 设备名称<br>Name of Equipment | 型号/编号<br>Model/Serial No. | 证书号/有效期<br>Certificate No./Due Date |
|---------------------------|---------------------------|-------------------------------------|
| 数字高压表                     | FRC-150M/11090603         | 2020DW02330030/2021-01-09           |

- 校准环境条件  
Environmental condition on the Calibration

|                           |                       |                                      |
|---------------------------|-----------------------|--------------------------------------|
| 温度: 20.1°C<br>Temperature | 相对湿度: 50 %<br>R.H.    | 其它: ____<br>Others                   |
| 气压: ____<br>Pressure      | 地点: 光谷基地B132<br>Place | 原始记录编号: 2020DW04400243<br>Record No. |

本校准结论, 仅对受校样品的本次校准有效。  
It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.  
未经本院许可, 不得部分复制本证书。

湖北省计量测试技术研究院  
Hubei Institute of Measurement and Testing Technology

证书编号: [2020DW04400243]号  
Certificate No.

**校准数据/结果**

Data/Results of Calibration

| 一、外观检查: 正常  |            |                           |
|-------------|------------|---------------------------|
| 二、高电压测量: AC |            |                           |
| 被测显示值 (kV)  | 实测标准值 (kV) | 测量不确定度<br>$U$ (kV), $k=2$ |
| 10.9        | 10.9       | 0.1                       |
| 20.4        | 20.4       | 0.2                       |
| 30.6        | 30.8       | 0.3                       |
| 40.5        | 40.6       | 0.4                       |
| 50.6        | 50.8       | 0.5                       |
| 60.2        | 60.4       | 0.6                       |

以下空白

