



广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO., LTD



中国合格评定  
国家认可委员会  
CALIBRATION  
CNAS L7117

# 校准证书

CALIBRATION CERTIFICATE



证书编号:

Certificate No.



扫一扫检验真伪

2GB19120517351-005

委托方:

Client

新乡市宝泉电力工程安装有限公司

委托方地址:

Address

河南省新乡市

仪器名称:

Description

雷击计数器校验仪

型号/规格:

Model/Type

FCZ-II

制造厂商:

Manufacturer

武汉三新电力设备制造有限公司

出厂编号:

Serial No

SXA1600

管理号:

Asset No.

样品接收日期:

Date of Receipt

2019-12-05

Y M D

结论:

Conclusion

所校准项目合格(Passed at Calibration Items)

校准日期:

Date of Calibration

2019-12-05

Y M D

建议下次校准日期:

Due Date

2020-12-04

Y M D

校准:

Calibration by

李健

审核:

Inspected by

侯利

批准人:

Approved Signatory

方文潮

(总经理)

证书专用章

(Stamp)

本实验室地址: 广东省广州市番禺区石碁镇农科所南街8号

Address: No.8 South Street Shi Ji Institute Guangzhou, Guangdong, China

联系电话(Tel): 020-31134076

传真(Fax): 020-31134076

公司网址(Web): www.lisaiest.com

客户电话(Tel): 020-31104772

邮政编码(Post): 511400

电子邮件(E-mail): cai@lisaiest.com



# 广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO., LTD

证书编号:

Certificate No.



中国合格评定  
国家认可  
校准  
CALIBRATION  
CNAS L1727

2GB19120517351-005

## 1、外观以及一般性检查:

In view of External and Generality check

正常

Pass

## 2、电压(Voltage):

| 指示值       | 标准值       | 误差    | 不确定度        | 允许误差 |
|-----------|-----------|-------|-------------|------|
| Indicated | Reference | Error | Uncertainty | MPE  |
| (V)       | (V)       | (V)   | (%)         | (V)  |
| 770       | 765       | 6     | 1.0         | N/A  |

附:

关于测量结果不确定度的说明:

appendix:

Directions of uncertainty in the calibration

1.依据(Reference document)

JF 1059.1-2012 测量不确定度评定与表示

(JF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2. 测量结果扩展不确定度 (Expanded uncertainty of the measurement results)

k=2

以下空白(The below is blank)

# LISAI



广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO., LTD



中国合格评定  
国家认可  
CALIBRATION  
CNAS 17117

# 校准说明

DIRECTIONS OF CALIBRATION

证书编号:

Certificate No.



2GB19120517351-005

1. 本机构是国家认可的计量检定机构, 质量管理体系符合ISO/IEC17025标准要求。

This station is the national authorized metrological verification station. The quality system is in accordance with ISO/IEC17025.

2. 本机构出具的数据均可溯源至国家计量基准和国际单位制(SI)。

All data issued by this station are traceable to national primary standards and International System of Units(SI)

3. 证书内页中“P”代表“Pass”, “F”代表“Fail”, “N/A”代表“Not Applicable”。

In the datasheet, “P” represents “Pass” and “F” represents “Fail” and “N/A” represents “Not Applicable”.

4. 环境条件。(Environmental condition during the calibration).

温度(Temperature): 22°C 相对湿度(Relative Humidity): 60 %

5. 校准地点。(Place of the Calibration).

客户现场(五楼实验室)

6. 被校准仪器限制使用条件。(Limiting condition of the instrument Calibration).

-----

7. 本次校准的技术依据。(Reference documents for the Calibration).

JJG 795-2016 耐电压测试仪检定规程 V. R. for Withstanding Voltage Testers

8. 本次校准使用的主要计量标准器具。(Main standards of measurement used in the Calibration).

| 名称<br>Description | 出厂编号<br>Serial No. | 证书号/有效期<br>Certificate No./ Due Date | 溯源机构<br>Traceability<br>Institute | 技术特征<br>Technique Character                                |
|-------------------|--------------------|--------------------------------------|-----------------------------------|------------------------------------------------------------|
| 数字高阻计             | 10061888           | 3GC1900001-0039/2020-01-01           | 广州力赛计量检测有限公司                      | DCV: ±(0.5%rdg + 0.03%range)<br>ACV: ±(1%rdg + 0.05%range) |

1. 本结果只与受检品/试验样品有关。The result relate only to the items calibrated/testing.

2. 未经本机构书面批准, 不得部分复制或全部复制。This certificate or report shall not be reproduced except in full, without the written approval of our station.

第2页 / 共3页

Page 2 of 3