

湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

中国校准服务联合体
China United Calibration Service

地址 (Add): 湖北省武汉市东湖新技术开发区茅

邮编 (Post Code): 430223

店山中路二号

电话 (Tel): 027-81925136

网址 (Web site): <http://www.hbjl.gov.cn>

传真 (Fax): 027-81925137

REGISTRATION NO. 002

校准证书

CALIBRATION CERTIFICATE

证书编号: [20190603570459]

Certificate No.

委托方名称 Customer	中国能源建设集团安徽电力建设第一工程有限公司
委托方地址 Address	安徽省合肥市
样品名称 Name Of Sample	绝缘电阻表
制造厂商 Manufacturer	武汉三新电力设备制造有限公司
型号规格 Model/Type	GM-5K
器具编号 No Of Sample	SXB3997

证书专用章

Stamp



校准日期
Calibration date

2019 年 06 月 20 日
Y M D

证书批准人
Approved by

核验员
Checked by

校准员
Calibrated by

耿睿
王强
吕夏

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准标准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



- 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据 (名称、代号)

Reference documents for the Calibration (Name, Code)

参照: JJG1005—2005 电子式绝缘电阻表检定规程 Verification Regulation of Electronic Insulating Resistance Meters

- 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称

Name of Equipment

绝缘电阻表检定装置

型号/编号

Model Serial No.

GZY92E/88843

证书号/有效期

Certificate No./Due Date

20180402250586/2019-09-10

- 校准环境条件

Environmental condition on the Calibration

温度: 20.8℃

Temperature

气压:

Pressure

相对湿度: 68 %

R.H.

地点: 本院光谷基地B211

Place

其它:

Others

原始记录编号: 2019003570459

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.
未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

电压选择(V)	500	1000	2500	5000	-----
开路电压(V)	507	1017	2547	5085	
$U_{rel} (k=2)$	1.4%	1.4%	1.4%	1.4%	
标准值(M Ω)	显示值				$U_{rel} (k=2)$
1.000	1.00 M Ω	1.00 M Ω	-----	-----	0.8%
2.000	2.00 M Ω	2.00 M Ω	-----	-----	
5.000	5.01 M Ω	5.00 M Ω	-----	-----	
10.00	10.0 M Ω	10.0 M Ω	10.0 M Ω	-----	
20.00	20.0 M Ω	20.0 M Ω	20.0 M Ω	-----	
50.00	49.9 M Ω	50.0 M Ω	50.1 M Ω	-----	1.2%
100.0	100 M Ω	100 M Ω	100 M Ω	100 M Ω	
200.0	200 M Ω	200 M Ω	200 M Ω	200 M Ω	
500.0	500 M Ω	500 M Ω	501 M Ω	501 M Ω	
1000	1.00 G Ω	1.00 G Ω	1.01 G Ω	1.01 G Ω	2.2%
2000	2.01 G Ω	2.01 G Ω	2.01 G Ω	2.00 G Ω	

以下空白